

# Certificate of Analysis

## HealthLink

3611 St. Johns Bluff Rd. So #1  
Jacksonville, FL 32224  
Phone: 904-996-7758  
Fax: 904-996-7078

## Dermatophyte Test Medium

Catalog # TR-7051

Lot # 0904021 Expiration 02-09-10

Physical Characteristics: Trace to slight hazy, yellow-orange  
pH @ 25°C: 5.5±0.2

Quality Assurance: Representative samples of this lot were tested with the following organisms under the conditions indicated:

Challenge Strain	ATCC #	Incubation	Results:
<i>Aspergillus niger</i>	16404	20-27°C up to 7 days	Inhibition
<i>Candida albicans</i>	10231	20-27°C up to 7 days	Growth
<i>Escherichia coli</i>	25922	20-27°C up to 7 days	Inhibition
<i>Trichophyton mentagrophytes</i>	9533	20-27°C up to 7 days	Good growth red color change

The Device Master Record for this lot was reviewed and approved to verify that this product was formulated and processed in full compliance with the product specifications.

This product was manufactured and tested according to the requirements stated in CFR 21 § 820.

By: Chris Timmons (signature on file)

Quality Assurance