

Certificate of Analysis

HealthLink

3611 St. Johns Bluff Rd. So #1
Jacksonville, FL 32224
Phone: 904-996-7758
Fax: 904-996-7078

Sabouraud Dextrose Agar

Catalog # TR-7050A

Lot # 0932313 Expiration 11-19-10

Physical Characteristics: Light amber, clear to very slightly opalescent without precipitate

pH @ 25°C: 5.7± 0.2

Quality Assurance: Representative samples of this lot were tested with the following organisms under the conditions indicated:

Challenge Strain	ATCC #	Incubation	Results:
<i>Aspergillus niger</i>	16404	25-30°C for up to 7 days	Good growth
<i>Candida albicans</i>	10231	25-30°C for up to 48 hrs.	Good growth
<i>Saccharomyces cerevisiae</i>	9763	25-30°C for up to 48 hrs.	Good growth

Microbial Load Analysis: A sample of this lot (based on ANSI sampling plan Z1.4) was tested for Microbial Load and passed.

The Device Master Record for this lot was reviewed and approved to verify that this product was formulated and processed in full compliance with the product specifications.

This product was manufactured and tested according to the requirements stated in CFR 21 § 820.

By: Chris Timmons (Signature on file)

Quality Assurance